

# EASYLIFE

## INSTRUCTIONS FOR CLEANING AND DISINFECTION

### Caution

- For Clinical Use Only.
- The following processes are validated for the EASYLIFE nasal mask when used in accordance with the manufacturer of the cleaning and disinfection instructions. Any deviation from the manufacturer's instructions or agents listed below may impact the performance or durability of the product. See cleaning and disinfection manufacturer's instructions for additional warnings and cautions.
- See the **Instructions for Use** included with the product for additional warnings and cautions.

### Warning

- Inspect the mask components after disinfection and replace any components that are worn or damaged. See the **Instructions for Use** included with the product for additional warnings and cautions.

### Notes

- The headgear cannot be processed using the cleaning and disinfection methods listed. The headgear must be replaced or hand/machine washed and allowed to air dry before reuse, as indicated in the **Instructions for Use** included with the product.
- After disinfection, discoloration of the cushion and a slight odor is normal.

### Cleaning

- Disassemble the mask per the **Instructions for Use** included with the product.
- Clean the mask using a soft bristle brush to adequately remove adhering substances from each component while soaking in a commercially available, anionic detergent such as MediZyme or Manu Klenz (used with chemical and gas plasma treatments). Extra attention should be given to the crevices and cavities during cleaning.

 NOTE: If compatible with the manufacturer's instructions, the detergent may be applied within the washer/disinfecter cycle.

- Rinse the mask with potable water and air dry out of direct sunlight. If cleaning manually, use a minimum 5 liters per mask.

### High-Level Disinfection

Mask and cushions may be reused after cleaning and disinfection by one of the chemical or gas plasma options below.

Chemical Disinfection: Apply only one treatment option following the manufacturer's instructions.

TREATMENT	EXPOSURE TIME	ACTIVE INGREDIENTS	SOLUTION	RINSE	MAXIMUM CYCLES
Cidex Activated Dialdehyde Solution	45 minutes	2.4% Glutaraldehyde	Full Strength	Potable Water, 5 Liters per Mask (minimum)	30
or					
Cidex OPA	12 minutes	0.55% Ortho-Phthalaldehyde	Full Strength	Potable Water, 5 Liters per Mask (minimum)	30
or					
Wavicide	45 minutes	2.65% Glutaraldehyde	Full Strength	Potable Water, 5 Liters per Mask (minimum)	30
or					
Control III	10 minutes	10% Alkyl Dimethyl Benzyl Ammonium Chloride  10% Alkyl Dimethyl Ethyl Benzyl Ammonium Chloride	8 ml per liter	Potable Water, 5 Liters per Mask (minimum)	30

## Hydrogen Peroxide Gas Plasma

TREATMENT	EXPOSURE TIME	MAXIMUM CYCLES
Sterrad 100S	Machine Controlled	30

Visit us on-line at [www.respironics.com](http://www.respironics.com) or contact customer service to obtain mask cleaning and disinfection instructions.

Respironics: 1-800-345-6443 (USA or Canada only) or 724-387-4000

Respironics Deutschland: +49 (0) 8152 93060

Respironics Inc.  
1001 Murry Ridge Lane  
Murrysville, PA 15668  
USA

Respironics Deutschland  
Gewerbestr. 17  
82211 Herrsching, Germany

EC REP

CE  
0123

Respironics is a registered trademark, and EASYLIFE is a trademark, respectively, of Respironics, Inc. and its affiliates. CIDEX, CIDEX OPA, and STERRAD are registered trademarks of the Advanced Sterilization Products. Wavicide is a registered trademark of Medical Chemical Corporation (MCC). Control III is a trademark of Maril Products, Inc. Manu Klenz is a trademark of Steris Corporation. MediZyme is a registered trademark of MEDICHEM International Ltd. ©2009 Respironics Inc. and its affiliates. All Rights Reserved.

1059850  
JAW 10/16/09